



Instructions for Catalog # 1030
Air and Emissions Lead in Impinger Solution
Revision 030512

Description:

- This standard is packaged in a 15 mL screw top vial containing approximately 14 mL of standard concentrate.
- This concentrate is preserved with approximately 2% (v/v) nitric acid.
- The concentrate can be stored at room temperature.
- The diluted standard will contain the Lead in the range specified on the data reporting form.

Before you begin:

- The sample resulting from the dilution described below will have a nitric acid concentration of approximately 0.02% before any acid is added. You may add a volume of acid different from the 4 to 10 mL of HNO₃ suggested in order to matrix match your calibration standards or meet any other method criteria.
- While it is technically not necessary to digest this standard prior to analysis, digestion should be performed if this is your normal procedure.
- This standard should be analyzed as soon as possible after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water and approximately 4 to 10 mL of nitric acid to a clean 1000 mL class A volumetric flask.
2. Shake the Lead in Impinger Solution vial prior to opening.
3. Using a clean, dry, class A pipet, volumetrically pipet 5.0 mL of the concentrate into the 1000 mL volumetric flask.
4. Dilute the flask to final volume with deionized water.
5. Cap the flask and mix well.
6. Immediately analyze the diluted sample by your normal procedures.
7. Report your results as µg/mL for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.