



Instructions for Catalog # 1151

Air and Emissions Particulate Matter in Impinger Solution

Revision 090119

Description:

- This standard is packaged in a 250 mL polyethylene bottle containing approximately 250 mL of standard.
- This standard is not preserved.
- The standard can be stored at room temperature.
- This product is intended to be used as a quality control check of the entire analytical process for the analytes/matrix included in the standard.
- ERA suggests that when subsampling this product prior to analysis you use a minimum sample size of at least 100 mL. Using a smaller sample size may invalidate the assigned value and/or uncertainty shown on the certificate of analysis.
- The certified values apply to the sample after following the stated instructions.

Helpful Hints:

- This standard has been prepared as a whole volume standard and is ready to analyze as received.
- Be sure to shake this sample very well prior to taking an aliquot for the Particulate Matter determination.

Instructions:

1. Shake the Particulate Matter in Impinger Solution bottle very well prior to opening.
2. Remove an aliquot and analyze by your normal procedures.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Safety Data Sheets (SDS) for all ERA products are available through our website www.eraqc.com.