



Instructions for Catalog # 1151QR
Air and Emissions Particulate Matter in Impinger Solution
Revision 071907

Description:

- This standard is packaged in a 250 mL polyethylene bottle containing approximately 250 mL of standard.
- This standard is not preserved.
- The standard can be stored at room temperature.
- The standard will contain Particulate Matter in the range of 140 to 675 mg/L.

Before you begin:

- This standard has been prepared as a whole volume standard and is ready to analyze as received.
- Be sure to shake this sample very well prior to taking an aliquot for the Particulate Matter determination.

Instructions:

1. Shake the Particulate Matter in Impinger Solution bottle very well prior to opening.
2. Remove an aliquot and analyze by your normal procedures.
3. Report your results as mg/L for the standard provided.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.