



Instructions for Catalog # 5270
WatR™ Supply Inorganic Disinfection #1
Revision 010312

Description:

- This standard is packaged in a 24 mL screw cap vial containing approximately 23 mL of standard concentrate.
- This concentrate is not preserved.
- The concentrate should be stored at $4\pm 2^{\circ}\text{C}$.
- The diluted standard will contain Chlorite and Chlorate in the ranges specified on the data reporting form.

Before you begin:

- This standard has been prepared as a concentrate and must be diluted prior to analysis.
- As the diluted standard is not stable, it must be analyzed immediately after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water to a clean 1000 mL class A volumetric flask.
2. Shake the Inorganic Disinfection #1 vial prior to opening.
3. Using a clean, dry, class A pipet, volumetrically pipet 5.0 mL of the concentrate into the 1000 mL volumetric flask.
4. Dilute the flask to final volume with deionized water.
5. Cap the flask and mix well.
6. Immediately analyze the diluted sample by your normal procedures.
7. Report your results as $\mu\text{g/L}$ for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.