



Instructions for Catalog # 551
WatR™ Supply Mercury
Revision 010312

Description:

- This standard is packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate.
- This concentrate is preserved with approximately 1% (v/v) nitric acid and 0.04% (w/v) potassium dichromate.
- The concentrate can be stored at room temperature.
- The diluted standard will contain Mercury in the range specified on the data reporting form.

Before you begin:

- This standard has been prepared as a concentrate and must be diluted prior to analysis.
- The USEPA *Criteria Document* requires that Mercury be present as a mixture of organic and inorganic forms and must, therefore, be analyzed as Total Mercury.
- This standard should be analyzed as soon as possible after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water and approximately 2 to 5 mL of nitric acid to a clean 500 mL class A volumetric flask.
2. Shake the Mercury vial prior to opening.
3. Using a clean, dry, class A pipet, volumetrically pipet 5.0 mL of the concentrate into the 500 mL volumetric flask.
4. Dilute the flask to final volume with deionized water.
5. Cap the flask and mix well.
6. Immediately analyze the diluted sample by your normal procedures.
7. Report your results as $\mu\text{g/L}$ for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.