



## Instructions for Catalog # 859

### WatR™ Supply Color

Revision 112414

#### Description:

- This standard is packaged in a 125 mL amber glass bottle containing approximately 125 mL of standard.
- This standard is preserved with approximately 1% (v/v) hydrochloric acid.
- This standard can be stored at room temperature.
- The standard has a Color in the range specified on the data reporting form.

#### Before you begin:

This standard has been prepared as a whole volume standard and should be analyzed as received. **DO NOT ADJUST THE pH OF THE SAMPLE.**

#### Instructions:

1. Shake the Color bottle prior to opening.
2. Remove an aliquot and analyze by your normal procedures.
3. Tightly recap the bottle immediately after removing an aliquot for analysis.
4. Report your results as PC color units for the standard provided.

#### Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.