



Instructions for Catalog # 900
WatR™ Supply Corrosivity
Revision 010312

Description:

- This standard is packaged in a 500 mL polyethylene bottle containing approximately 500 mL of standard.
- This standard is not preserved.
- The standard can be stored at room temperature.
- This standard contains Corrosivity in the range specified on the data reporting form.

Before you begin:

- This standard has been prepared as a whole volume standard and is ready to be analyzed as received.
- If the Corrosivity standard is to be opened and resealed over a period of time, make sure to tightly cap the bottle immediately after each use.

Instructions:

1. Shake the Corrosivity bottle prior to opening.
2. Remove an aliquot and analyze by your normal procedures.
3. Tightly recap the bottle after each use.
4. Report your results as saturation index units @ 20°C for the standard provided.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.