



Instructions for Catalog # 980 WatR™ Supply Corrosivity

Revision 090119

Description:

- This standard is packaged in a 500 mL polyethylene bottle containing approximately 500 mL of standard.
- This standard is not preserved.
- The standard can be stored at room temperature.
- This product is intended to be used as a quality control check of the entire analytical process for the analytes/matrix included in the standard.
- ERA suggests that when subsampling this product prior to analysis you use a minimum sample size of at least 10 mL. Using a smaller sample size may invalidate the assigned value and/or uncertainty shown on the certificate of analysis.
- Repeated sampling of this product is permitted, provided minimum sample sizes and storage instructions are adhered to.
- The certified values apply to the sample after following the stated instructions.

Helpful Hints:

- This standard has been prepared as a whole volume standard and is ready to be analyzed as received.
- If the Corrosivity standard is to be opened and resealed over a period of time, make sure to tightly cap the bottle immediately after each use.

Instructions:

1. Shake the Corrosivity bottle prior to opening.
2. Remove an aliquot and analyze by your normal procedures.
3. Tightly recap the bottle after each use.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Safety Data Sheets (SDS) for all ERA products are available through our website. www.eraqc.com