



Instructions for Catalog # 4990  
WatR™ Pollution Lithium  
Revision 080411

Description:

- This standard is packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate.
- This concentrate is preserved with approximately 1% nitric acid.
- The concentrate can be stored at room temperature.
- The diluted standard will contain Lithium in the range of 50 to 500 µg/L.

Before you begin:

- This standard is designed for the Ohio VAP program.
- This standard has been prepared as a concentrate and must be diluted prior to analysis.
- This standard should be analyzed as soon as possible after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water and 2-5 mL of nitric acid to a clean 1000 mL class A volumetric flask.
2. Shake the Lithium vial prior to opening.
3. Using a clean, dry, class A pipet, volumetrically pipet 10.0 mL of the concentrate into the 1000 mL volumetric flask.
4. Dilute the flask to final volume with deionized water.
5. Cap the flask and mix well.
6. Immediately analyze the diluted sample by your normal procedures.
7. Report your results as µg/L for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.