



Instructions for Catalog # 499QR  
WatR™ Pollution Solids  
Revision 030512

Description:

- This standard is packaged in a 500 mL polyethylene bottle containing approximately 500 mL of standard.
- This standard is not preserved.
- The standard can be stored at room temperature.
- The standard will contain all the analytes listed in the ranges specified on the data reporting form.

Before you begin:

- This standard has been prepared as a whole volume standard and is ready to analyze as received.
- Be sure to shake this sample very well prior to taking an aliquot for the TSS and TS determinations.
- It is strongly recommended that you analyze for TSS prior to removing aliquots for other analyses from the Solids bottle.
- If other analytes are to be measured prior to TSS, be sure to shake the sample very well prior to removing aliquots for any of the other parameters.

Instructions:

1. Shake the Solids bottle very well prior to opening.
2. Remove aliquots for the tests that you are performing and analyze by your normal procedures.
3. Report your results as mg/L for the standard provided.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.