



Instructions for Catalog # 911QR WatR™ Pollution Settleable Solids

Revision 040307

Description:

- This standard is packaged in a 60 mL polyethylene bottle containing a solid concentrate.
- This concentrate is not preserved.
- The concentrate can be stored at room temperature.
- The diluted standard will contain Settleable Solids in the range of 5 to 100 mL/L.

Before you begin:

- This standard has been prepared as a concentrate and must be diluted prior to analysis.
- This standard should be analyzed as soon as possible after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water to a clean 1000 mL class A volumetric flask.
2. Transfer the contents of the Settleable Solids bottle to the 1000 mL flask.
3. Rinse the Settleable Solids bottle and cap three times with deionized water, adding the rinsate to the flask each time.
4. Dilute the flask to final volume with deionized water.
5. Mix the sample with a stir bar for ten minutes.
6. Transfer the entire 1 liter sample to an Imhoff cone and proceed per your normal procedures.
7. Report your results as mL/L for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.